




Data File : VV012064.D
Acq On : 29 Jul 2019 15:59
Operator : SY/MD
Sample : MDL02
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Jul 30 08:50:11 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR072919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 30 08:46:06 2019
Response via : Initial Calibration



TIC: VV012064.D

(21) 2-Butanone (T)
4.013min (-0.035) 0.13ug/L

response 257

Ion	Exp%	Act%
43.05	100	100
72.10	26.80	39.30
0.00	0.00	0.00
0.00	0.00	0.00