

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072922\
 Data File : VW027095.D
 Acq On : 29 Jul 2022 19:22
 Operator : SY/MD
 Sample : N3899-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUR1

Quant Time: Jul 30 00:10:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 30 00:04:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	555881	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	546038	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	245607	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	218243	41.268	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.540%
7) Chloroethane-d5	1.564	69	184506	44.033	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.060%
11) 1,1-Dichloroethene-d2	2.101	63	299776	33.296	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.600%
21) 2-Butanone-d5	3.886	46	191990	63.667	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	63.670%
24) Chloroform-d	4.342	84	401724	42.176	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.360%
26) 1,2-Dichloroethane-d4	5.027	65	237527	43.930	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.860%
32) Benzene-d6	5.043	84	862906	45.002	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.000%
36) 1,2-Dichloropropane-d6	6.066	67	269727	45.959	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.920%
41) Toluene-d8	7.313	98	712444	41.107	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.220%
43) trans-1,3-Dichloroprop...	7.619	79	103852	40.059	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.120%
47) 2-Hexanone-d5	8.088	63	202612	80.307	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.310%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	371026	41.929	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.860%
66) 1,2-Dichlorobenzene-d4	11.622	152	260267	47.140	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072922\
Data File : VV027095.D
Acq On : 29 Jul 2022 19:22
Operator : SY/MD
Sample : N3899-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUR1

Quant Time: Jul 30 00:10:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 30 00:04:45 2022
Response via : Initial Calibration

