

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072922\
 Data File : VW027113.D
 Acq On : 30 Jul 2022 03:32
 Operator : SY/MD
 Sample : N3899-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUR7

Quant Time: Jul 30 04:30:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 30 01:28:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	674180	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	639058	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	284068	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	230136	35.881	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.760%
7) Chloroethane-d5	1.564	69	190121	37.412	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.820%
11) 1,1-Dichloroethene-d2	2.101	63	309095	28.307	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.620%#
21) 2-Butanone-d5	3.892	46	245757	67.197	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	67.200%
24) Chloroform-d	4.342	84	449576	38.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.840%
26) 1,2-Dichloroethane-d4	5.027	65	268334	40.919	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.840%
32) Benzene-d6	5.043	84	955597	42.582	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.160%
36) 1,2-Dichloropropane-d6	6.066	67	295434	43.012	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.020%
41) Toluene-d8	7.313	98	807716	39.820	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.640%#
43) trans-1,3-Dichloroprop...	7.619	79	119981	39.544	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.080%
47) 2-Hexanone-d5	8.088	63	232945	78.891	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.890%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	405683	39.173	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.340%
66) 1,2-Dichlorobenzene-d4	11.622	152	290960	45.564	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.120%
Target Compounds						
					Qvalue	
33) Benzene	5.098	78	26827	1.303	ug/L	100
51) Chlorobenzene	8.882	112	83523	5.833	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072922\
Data File : VV027113.D
Acq On : 30 Jul 2022 03:32
Operator : SY/MD
Sample : N3899-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUR7

Quant Time: Jul 30 04:30:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 30 01:28:32 2022
Response via : Initial Calibration

