

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW072922\  
 Data File : VW027118.D  
 Acq On : 30 Jul 2022 05:34  
 Operator : SY/MD  
 Sample : N3899-09  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 48 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 DBUR6

Quant Time: Jul 30 06:09:44 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM072822WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Sat Jul 30 01:28:32 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	602247	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	589454	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	254285	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	221412	38.644	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery =	77.280%		
7) Chloroethane-d5	1.564	69	186106	40.996	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery =	82.000%		
11) 1,1-Dichloroethene-d2	2.101	63	308253	31.602	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	63.200%		
21) 2-Butanone-d5	3.886	46	200766	61.452	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery =	61.450%		
24) Chloroform-d	4.342	84	424362	41.123	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	82.240%		
26) 1,2-Dichloroethane-d4	5.027	65	248879	42.486	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	84.980%		
32) Benzene-d6	5.043	84	896594	43.315	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	86.620%		
36) 1,2-Dichloropropane-d6	6.066	67	273916	43.235	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery =	86.480%		
41) Toluene-d8	7.313	98	759333	40.585	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	81.180%		
43) trans-1,3-Dichloroprop...	7.619	79	104446	37.320	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	74.640%		
47) 2-Hexanone-d5	8.088	63	212653	78.079	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery =	78.080%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	381918	39.981	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery =	79.960%		
66) 1,2-Dichlorobenzene-d4	11.622	152	267139	46.733	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	93.460%		
Target Compounds						
51) Chlorobenzene	8.879	112	11454	0.867	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV072922\  
 Data File : VV027118.D  
 Acq On : 30 Jul 2022 05:34  
 Operator : SY/MD  
 Sample : N3899-09  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 48 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 DBUR6

Quant Time: Jul 30 06:09:44 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM072822WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Sat Jul 30 01:28:32 2022  
 Response via : Initial Calibration

