

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072922\
 Data File : VW027080.D
 Acq On : 29 Jul 2022 13:14
 Operator : SY/MD
 Sample : N3841-18
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jul 30 00:06:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 30 00:04:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	531278	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	540668	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	227478	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	242524	47.983	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	95.960%	
7) Chloroethane-d5	1.564	69	208035	51.948	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	103.900%	
11) 1,1-Dichloroethene-d2	2.101	63	336547	39.111	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	78.220%	
21) 2-Butanone-d5	3.892	46	225335	78.185	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	78.190%	
24) Chloroform-d	4.342	84	464673	51.045	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.080%	
26) 1,2-Dichloroethane-d4	5.027	65	282496	54.666	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	109.340%	
32) Benzene-d6	5.046	84	955873	50.345	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.700%	
36) 1,2-Dichloropropane-d6	6.066	67	297072	51.121	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	102.240%	
41) Toluene-d8	7.313	98	781455	45.536	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.080%	
43) trans-1,3-Dichloroprop...	7.619	79	119776	46.660	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.320%	
47) 2-Hexanone-d5	8.088	63	225145	90.125	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	90.120%	
56) 1,1,2,2-Tetrachloroeth...	10.213	84	428135	48.864	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	97.720%	
66) 1,2-Dichlorobenzene-d4	11.622	152	284091	55.555	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	111.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072922\
Data File : VV027080.D
Acq On : 29 Jul 2022 13:14
Operator : SY/MD
Sample : N3841-18
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jul 30 00:06:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 30 00:04:45 2022
Response via : Initial Calibration

