

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW072922\
 Data File : VW027097.D
 Acq On : 29 Jul 2022 20:11
 Operator : SY/MD
 Sample : N3899-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUS2

Quant Time: Jul 30 00:10:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 30 00:04:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	562324	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	553499	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	234476	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	214127	40.025	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.060%
7) Chloroethane-d5	1.564	69	185292	43.714	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.420%
11) 1,1-Dichloroethene-d2	2.101	63	295522	32.447	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.900%
21) 2-Butanone-d5	3.889	46	204477	67.031	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	67.030%
24) Chloroform-d	4.342	84	412762	42.839	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.680%
26) 1,2-Dichloroethane-d4	5.027	65	249394	45.596	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.200%
32) Benzene-d6	5.043	84	857412	44.113	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.220%
36) 1,2-Dichloropropane-d6	6.066	67	271795	45.687	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.380%
41) Toluene-d8	7.313	98	707091	40.248	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.500%
43) trans-1,3-Dichloroprop...	7.619	79	103687	39.456	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.920%
47) 2-Hexanone-d5	8.088	63	202176	79.054	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.050%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	374790	41.784	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.560%
66) 1,2-Dichlorobenzene-d4	11.622	152	249570	47.348	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV072922\
Data File : VV027097.D
Acq On : 29 Jul 2022 20:11
Operator : SY/MD
Sample : N3899-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUS2

Quant Time: Jul 30 00:10:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM072822WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 30 00:04:45 2022
Response via : Initial Calibration

