

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073018\
 Data File : VV006756.D
 Acq On : 30 Jul 2018 11:47
 Operator : SY/MD
 Sample : VV0730WBL01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK74

Quant Time: Aug 01 08:48:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 28 03:46:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	292213	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	266012	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	106092	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	84936	5.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.80%
7) Chloroethane-d5	1.58	69	66278	5.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.00%
11) 1,1-Dichloroethene-d2	2.13	63	117318	3.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.60%
20) 2-Butanone-d5	3.95	46	185832	39.11	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.22%
24) Chloroform-d	4.41	84	137782	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	5.09	65	68619	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.10	84	299548	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.60%
36) 1,2-Dichloropropane-d6	6.12	67	93467	5.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.20%
41) Toluene-d8	7.36	98	257845	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.80%
43) trans-1,3-Dichloropropene-	7.67	79	30239	5.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.40%
46) 2-Hexanone-d5	8.14	63	141193	37.00	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.00%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	61825	4.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	83644	5.29	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073018\
Data File : VV006756.D
Acq On : 30 Jul 2018 11:47
Operator : SY/MD
Sample : VV0730WBL01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK74

Quant Time: Aug 01 08:48:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 28 03:46:17 2018
Response via : Initial Calibration

