

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073018\
 Data File : VV006768.D
 Acq On : 30 Jul 2018 17:49
 Operator : SY/MD
 Sample : J4207-18
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jul 30 18:13:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jul 30 16:58:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	188129	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	176487	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	68513	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	52865	5.40	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.00%
7) Chloroethane-d5	1.58	69	44054	5.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.60%
11) 1,1-Dichloroethene-d2	2.13	63	73046	3.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.00%
20) 2-Butanone-d5	3.95	46	164169	53.66	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.32%
24) Chloroform-d	4.40	84	100197	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
26) 1,2-Dichloroethane-d4	5.09	65	52110	5.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.40%
32) Benzene-d6	5.10	84	202803	5.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.80%
36) 1,2-Dichloropropane-d6	6.12	67	68686	5.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.00%
41) Toluene-d8	7.36	98	165554	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
43) trans-1,3-Dichloropropene-	7.67	79	20030	5.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.20%
46) 2-Hexanone-d5	8.14	63	118567	46.83	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.66%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	52101	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	60142	5.89	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.80%

Target Compounds					Ovalue
13) Acetone	2.19	43	5981	3.07 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073018\
Data File : VV006768.D
Acq On : 30 Jul 2018 17:49
Operator : SY/MD
Sample : J4207-18
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jul 30 18:13:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jul 30 16:58:50 2018
Response via : Initial Calibration

