


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073018\
 Data File : VV006754.D
 Acq On : 30 Jul 2018 09:34
 Operator : SY/MD
 Sample : BFB73
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB73

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Aug 01 01:43:21 2018



AutoFind: Scans 2871, 2872, 2873; Background Corrected with Scan 2858

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	11025	PASS
75	95	30	80	47.4	30714	PASS
95	95	100	100	100.0	64848	PASS
96	95	5	9	6.4	4146	PASS
173	174	0.00	2	0.6	320	PASS
174	95	50	120	81.4	52778	PASS
175	174	5	9	7.1	3773	PASS
176	174	95	101	97.0	51221	PASS
177	176	5	9	6.5	3346	PASS

Scan 2858 (10.080 min): VV006754.D
BFB73

m/z	Abundance
39.90	246.0
43.90	626.0

Instrument :
MSVOA_V
ClientSampleId :
BFB73

m/z	Abundance
35.90	459.0
36.90	2444.0
37.90	2472.0
38.90	1007.0
39.90	235.0
43.80	866.0
44.80	478.0
45.80	108.0
46.90	855.0
47.80	393.0
48.90	2315.0
49.90	10709.0
50.90	3482.0
51.90	159.0
54.80	175.0
55.90	877.0
56.90	1473.0
59.90	576.0
60.90	2888.0
61.90	2704.0
62.90	2141.0
63.80	262.0
64.90	496.0
66.90	170.0
67.90	5678.0
68.90	6117.0
70.00	494.0
71.90	341.0
72.90	2693.0
73.90	9512.0
74.90	29624.0
75.90	2704.0
76.90	467.0
77.80	270.0
78.80	1583.0
79.80	542.0
80.80	1613.0
81.90	411.0
86.80	2668.0
87.80	2770.0
90.90	216.0
91.90	1660.0
92.90	2507.0
93.90	6667.0
94.90	61880.0
95.90	4121.0
97.00	155.0
104.00	201.0
105.80	250.0
115.80	179.0
116.70	367.0
117.60	228.0
118.80	282.0
127.80	209.0
128.70	105.0
129.80	278.0
140.80	467.0
142.80	539.0
144.80	131.0
154.70	145.0
171.80	642.0
173.00	479.0
173.80	49200.0
174.90	3428.0
175.80	47200.0
176.80	3118.0
177.80	110.0

Instrument :
MSVOA_V
ClientSampleId :
BFB73

m/z	Abundance
35.90	464.0
36.90	2505.0
37.90	2685.0
38.90	1069.0
39.90	254.0
43.80	817.0
44.90	520.0
45.90	114.0
46.90	928.0
47.80	405.0
48.90	2490.0
49.90	11396.0
50.90	3713.0
51.90	173.0
54.80	175.0
55.90	860.0
56.90	1606.0
59.90	602.0
60.90	3042.0
61.90	2934.0
62.90	2276.0
63.80	291.0
64.90	573.0
66.90	198.0
67.90	6165.0
68.90	6460.0
69.90	512.0
71.90	315.0
72.90	2829.0
73.90	10238.0
74.90	31688.0
75.90	2786.0
76.90	480.0
77.80	274.0
78.80	1687.0
79.90	599.0
80.80	1697.0
81.90	411.0
86.00	105.0
86.90	2965.0
87.80	2993.0
90.80	224.0
91.90	1789.0
92.90	2716.0
93.90	7373.0
94.90	66960.0
95.90	4239.0
96.90	192.0
104.00	218.0
105.80	257.0
115.80	181.0
116.70	400.0
117.60	258.0
118.80	327.0
127.80	227.0
128.70	118.0
129.80	268.0
140.80	529.0
142.80	602.0
144.90	160.0
147.80	111.0
154.70	152.0
171.80	720.0
173.80	54520.0
174.90	3904.0
175.80	52912.0
176.90	3473.0

Instrument :
MSVOA_V
ClientSampleId :
BFB73

m/z	Abundance
35.90	481.0
36.90	2453.0
37.90	2577.0
38.90	1103.0
39.90	286.0
40.90	108.0
43.80	734.0
44.90	530.0
46.00	116.0
46.80	895.0
47.80	380.0
48.90	2423.0
49.90	10970.0
50.90	3611.0
51.90	163.0
54.80	163.0
55.90	774.0
56.90	1573.0
59.90	556.0
60.90	2885.0
61.90	2862.0
62.90	2154.0
63.90	317.0
64.90	648.0
66.90	208.0
67.90	6206.0
68.90	6188.0
69.90	503.0
71.90	267.0
72.90	2703.0
73.90	10102.0
74.90	30832.0
75.90	2665.0
76.90	441.0
77.80	255.0
78.80	1658.0
79.90	606.0
80.90	1636.0
81.80	396.0
86.00	119.0
86.90	3046.0
87.80	2889.0
90.80	219.0
91.90	1700.0
92.90	2689.0
93.90	7305.0
94.90	65704.0
95.90	4078.0
96.90	218.0
103.90	191.0
105.80	239.0
111.80	107.0
115.80	215.0
116.80	379.0
117.70	236.0
118.80	332.0
127.70	212.0
128.70	117.0
129.80	233.0
134.70	106.0
140.80	520.0
142.80	622.0
144.80	180.0
145.70	111.0
147.80	127.0
154.70	136.0
156.60	101.0
171.90	749.0
173.00	482.0
173.80	54616.0
174.90	3989.0
175.80	53552.0
176.90	3448.0

Instrument :
MSVOA_V
ClientSampleId :
BFB73