

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073019\
 Data File : VV012068.D
 Acq On : 30 Jul 2019 11:02
 Operator : SY/MD
 Sample : VV0730WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK59

Quant Time: Jul 31 09:30:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR072919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 31 09:22:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	134499	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	127847	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.30	152	62592	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	30734	4.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.80%
7) Chloroethane-d5	1.59	69	24430	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.60%
11) 1,1-Dichloroethene-d2	2.13	63	48836	3.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.80%
20) 2-Butanone-d5	3.97	46	82092	47.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.98%
24) Chloroform-d	4.40	84	78810	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.08	65	43942	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	5.10	84	154638	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.12	67	43947	4.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.60%
41) Toluene-d8	7.36	98	142913	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	7.67	79	17754	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
46) 2-Hexanone-d5	8.14	63	62344	44.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.54%
56) 1,1,2,2-Tetrachloroethane-	10.26	84	31172	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
66) 1,2-Dichlorobenzene-d4	11.67	152	55645	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073019\
Data File : VV012068.D
Acq On : 30 Jul 2019 11:02
Operator : SY/MD
Sample : VV0730WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK59

Quant Time: Jul 31 09:30:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR072919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 31 09:22:51 2019
Response via : Initial Calibration

