

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073020\
 Data File : VV017756.D
 Acq On : 30 Jul 2020 18:42
 Operator : SY/MD
 Sample : L3388-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleID :
 C0AA1

Quant Time: Jul 31 08:04:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 31 02:37:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	300712	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	290061	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	140333	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	58528	39.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.82%
7) Chloroethane-d5	1.58	69	55985	44.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.42%
11) 1,1-Dichloroethene-d2	2.12	63	107756	30.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.16%
21) 2-Butanone-d5	3.92	46	120209	109.02	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.02%
24) Chloroform-d	4.37	84	190156	47.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.90%
26) 1,2-Dichloroethane-d4	5.06	65	146132	49.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.84%
32) Benzene-d6	5.07	84	330672	48.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.36%
36) 1,2-Dichloropropane-d6	6.09	67	96003	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.30%
41) Toluene-d8	7.34	98	306506	45.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.10%
43) trans-1,3-Dichloropropene-	7.64	79	55412	44.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.48%
47) 2-Hexanone-d5	8.11	63	80743	99.12	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.12%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	134817	47.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	139101	50.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.20	43	14542	10.385	ug/L	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073020\
 Data File : VV017756.D
 Acq On : 30 Jul 2020 18:42
 Operator : SY/MD
 Sample : L3388-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 C0AA1

Quant Time: Jul 31 08:04:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 31 02:37:22 2020
 Response via : Initial Calibration

