

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073020\
 Data File : VV017760.D
 Acq On : 30 Jul 2020 21:02
 Operator : SY/MD
 Sample : L3486-05
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0Y28

Quant Time: Jul 31 08:14:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 31 02:37:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	316705	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	303434	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	147408	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	68023	44.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.10%
7) Chloroethane-d5	1.58	69	60700	46.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.06%
11) 1,1-Dichloroethene-d2	2.12	63	124925	33.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.32%
21) 2-Butanone-d5	3.91	46	112112	96.54	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.54%
24) Chloroform-d	4.37	84	192262	45.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.10%
26) 1,2-Dichloroethane-d4	5.06	65	148849	48.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.56%
32) Benzene-d6	5.07	84	341033	47.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.09	67	97711	49.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.56%
41) Toluene-d8	7.34	98	323730	45.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.98%
43) trans-1,3-Dichloropropene-	7.64	79	57889	44.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.36%
47) 2-Hexanone-d5	8.11	63	78858	92.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.54%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	132407	44.23	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.46%
64) 1,2-Dichlorobenzene-d4	11.65	152	146026	50.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.94%

Target Compounds					Ovalue
9) Trichlorofluoromethane	1.76	101	8409	1.906 ug/L	96
34) Trichloroethene	5.94	95	5576	2.384 ug/L	79
46) Tetrachloroethene	8.00	164	20743	10.009 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073020\
Data File : VV017760.D
Acq On : 30 Jul 2020 21:02
Operator : SY/MD
Sample : L3486-05
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0Y28

Quant Time: Jul 31 08:14:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 31 02:37:22 2020
Response via : Initial Calibration

