

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073020\
 Data File : VV017747.D
 Acq On : 30 Jul 2020 13:40
 Operator : SY/MD
 Sample : VV0730WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK65

Quant Time: Jul 31 02:55:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 31 02:37:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	340828	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	317605	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	156322	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80109	48.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.40%
7) Chloroethane-d5	1.58	69	67132	47.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.62%
11) 1,1-Dichloroethene-d2	2.12	63	143636	35.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.92%
21) 2-Butanone-d5	3.91	46	121973	97.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.60%
24) Chloroform-d	4.37	84	212674	46.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.64%
26) 1,2-Dichloroethane-d4	5.05	65	161830	48.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.56%
32) Benzene-d6	5.07	84	379840	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.10%
36) 1,2-Dichloropropane-d6	6.09	67	107454	51.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.56%
41) Toluene-d8	7.34	98	360886	48.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.96%
43) trans-1,3-Dichloropropene-	7.64	79	63313	46.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.32%
47) 2-Hexanone-d5	8.11	63	84503	94.74	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.74%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	143019	45.64	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.28%
64) 1,2-Dichlorobenzene-d4	11.65	152	158552	51.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.34%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073020\
Data File : VV017747.D
Acq On : 30 Jul 2020 13:40
Operator : SY/MD
Sample : VV0730WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK65

Quant Time: Jul 31 02:55:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 31 02:37:22 2020
Response via : Initial Calibration

