

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073118\
 Data File : VV006777.D
 Acq On : 31 Jul 2018 10:43
 Operator : SY/MD
 Sample : VV0731WBL01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK75

Quant Time: Aug 01 01:51:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 01 01:43:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	177034	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	163825	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	65733	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48229	5.23	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.60%
7) Chloroethane-d5	1.58	69	43236	5.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.60%
11) 1,1-Dichloroethene-d2	2.13	63	68161	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	3.94	46	161797	56.20	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.40%
24) Chloroform-d	4.40	84	93022	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.60%
26) 1,2-Dichloroethane-d4	5.09	65	51277	6.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.60%
32) Benzene-d6	5.10	84	190702	5.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.20%
36) 1,2-Dichloropropane-d6	6.12	67	65111	5.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.40%
41) Toluene-d8	7.36	98	152144	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
43) trans-1,3-Dichloropropene-	7.67	79	19587	5.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.60%
46) 2-Hexanone-d5	8.14	63	115650	49.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.40%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	50957	5.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	55642	5.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073118\
Data File : VV006777.D
Acq On : 31 Jul 2018 10:43
Operator : SY/MD
Sample : VV0731WBL01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK75

Quant Time: Aug 01 01:51:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 01 01:43:21 2018
Response via : Initial Calibration

