

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073118\
 Data File : VV006780.D
 Acq On : 31 Jul 2018 12:18
 Operator : SY/MD
 Sample : J4225-07
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWP4

Quant Time: Aug 01 02:01:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 01 01:43:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	172378	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	161270	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	63062	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	44412	4.95	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.00%
7) Chloroethane-d5	1.58	69	41325	5.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.40%
11) 1,1-Dichloroethene-d2	2.13	63	65044	3.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.00%
20) 2-Butanone-d5	3.95	46	145178	51.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.58%
24) Chloroform-d	4.40	84	93720	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.40%
26) 1,2-Dichloroethane-d4	5.09	65	48797	5.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.00%
32) Benzene-d6	5.10	84	184739	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.40%
36) 1,2-Dichloropropane-d6	6.12	67	63412	5.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.20%
41) Toluene-d8	7.36	98	146828	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
43) trans-1,3-Dichloropropene-	7.67	79	18598	5.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.80%
46) 2-Hexanone-d5	8.14	63	102802	44.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.86%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	47818	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	54924	5.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.80%

Target Compounds					Ovalue
13) Acetone	2.19	43	10811	6.06	ug/L 96
33) Benzene	5.15	78	20826	0.43	ug/L 100
42) Toluene	7.44	91	28604	0.57	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073118\
Data File : VV006780.D
Acq On : 31 Jul 2018 12:18
Operator : SY/MD
Sample : J4225-07
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWP4

Quant Time: Aug 01 02:01:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 01 01:43:21 2018
Response via : Initial Calibration

