

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073118\
 Data File : VV006785.D
 Acq On : 31 Jul 2018 14:58
 Operator : SY/MD
 Sample : J4225-16
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWR2

Quant Time: Aug 01 02:27:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 01 01:43:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	202671	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	188664	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	79397	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	46524	4.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.20%
7) Chloroethane-d5	1.58	69	46769	5.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.00%
11) 1,1-Dichloroethene-d2	2.13	63	68885	3.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.60%
20) 2-Butanone-d5	3.95	46	165051	50.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.16%
24) Chloroform-d	4.41	84	106817	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.00%
26) 1,2-Dichloroethane-d4	5.09	65	54650	5.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.20%
32) Benzene-d6	5.10	84	209051	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.80%
36) 1,2-Dichloropropane-d6	6.12	67	72499	5.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.60%
41) Toluene-d8	7.36	98	170454	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
43) trans-1,3-Dichloropropene-	7.67	79	21248	4.97	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.40%
46) 2-Hexanone-d5	8.14	63	133007	49.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.28%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	56467	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	66168	5.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.80%

Target Compounds					Ovalue
3) Chloromethane	1.26	50	10384	0.67 ug/L	89
13) Acetone	2.19	43	6687	3.19 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073118\
Data File : VV006785.D
Acq On : 31 Jul 2018 14:58
Operator : SY/MD
Sample : J4225-16
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWR2

Quant Time: Aug 01 02:27:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 01 01:43:21 2018
Response via : Initial Calibration

