


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073118\
Data File : VV006775.D
Acq On : 31 Jul 2018 09:06
Operator : SY/MD
Sample : BFB74
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB74

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072318WMA.M
Title : TRACE VOA SOM01.0
Last Update : Wed Aug 01 01:43:21 2018



AutoFind: Scans 2871, 2872, 2873; Background Corrected with Scan 2858

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	13686	PASS
75	95	30	80	48.1	37338	PASS
95	95	100	100	100.0	77701	PASS
96	95	5	9	6.7	5234	PASS
173	174	0.00	2	1.0	610	PASS
174	95	50	120	81.2	63090	PASS
175	174	5	9	7.1	4488	PASS
176	174	95	101	96.3	60765	PASS
177	176	5	9	6.5	3942	PASS

Scan 2858 (10.080 min): VV006775.D
BFB74

m/z	Abundance
39.90	192.0
43.80	520.0

Instrument :
MSVOA_V
ClientSampleId :
BFB74

m/z	Abundance
35.90	543.0
36.90	2915.0
37.90	2900.0
39.00	1226.0
39.80	249.0
43.90	817.0
44.90	569.0
46.00	151.0
46.90	961.0
47.80	464.0
48.90	2874.0
49.90	13576.0
50.90	4178.0
51.90	163.0
52.80	123.0
54.90	198.0
55.90	1011.0
56.80	1984.0
57.90	136.0
59.80	662.0
60.90	3187.0
61.90	3585.0
62.90	2811.0
63.80	392.0
65.00	612.0
66.70	218.0
67.90	7170.0
68.90	7389.0
70.00	537.0
71.90	365.0
72.90	3208.0
73.90	11681.0
74.90	36464.0
75.90	3385.0
76.90	533.0
78.80	1990.0
79.90	688.0
80.80	1854.0
81.80	444.0
85.80	103.0
86.90	3488.0
87.90	3460.0
90.70	280.0
91.90	2136.0
92.90	3209.0
93.90	8821.0
94.90	75496.0
95.90	5154.0
97.10	163.0
103.70	312.0
105.90	297.0
115.80	248.0
116.80	481.0
117.80	212.0
118.80	388.0
127.80	245.0
128.70	118.0
129.80	202.0
134.70	138.0
136.80	101.0
140.80	667.0
142.90	664.0
145.00	160.0
145.70	101.0
147.70	160.0
154.80	157.0
156.80	143.0
171.80	780.0
172.90	568.0
173.80	59400.0
174.90	4359.0
175.80	57216.0
176.90	3818.0
177.80	106.0

Instrument :
MSVOA_V
ClientSampleId :
BFB74

m/z	Abundance
35.90	597.0
36.90	3107.0
37.90	3029.0
38.90	1285.0
39.80	233.0
43.90	860.0
44.90	601.0
46.10	123.0
46.90	991.0
47.80	508.0
48.90	3018.0
49.90	14203.0
50.90	4397.0
51.90	183.0
52.80	111.0
54.90	207.0
55.90	1073.0
56.80	2071.0
57.90	129.0
59.80	714.0
60.90	3218.0
61.90	3650.0
62.90	2977.0
63.80	402.0
64.90	688.0
66.70	231.0
67.90	7570.0
68.90	7733.0
69.90	565.0
71.90	411.0
72.90	3328.0
73.90	12445.0
74.90	38632.0
75.90	3442.0
76.90	655.0
78.80	2012.0
79.80	731.0
80.80	1981.0
81.80	473.0
85.80	123.0
86.90	3739.0
87.90	3861.0
90.70	293.0
91.90	2331.0
92.90	3437.0
93.90	9284.0
94.90	80144.0
95.90	5389.0
97.10	170.0
103.70	285.0
104.90	108.0
105.80	326.0
106.80	112.0
115.80	259.0
116.80	533.0
117.70	276.0
118.80	435.0
127.80	268.0
128.70	121.0
129.70	215.0
130.70	107.0
134.80	143.0
136.70	118.0
140.80	697.0
142.90	790.0
144.70	154.0
145.00	170.0
145.80	113.0
147.70	184.0
154.80	177.0
156.80	130.0
171.80	916.0
172.90	627.0
173.80	65200.0
174.90	4642.0
175.80	62888.0
176.90	4069.0
177.80	124.0

Instrument :
MSVOA_V
ClientSampleId :
BFB74

m/z	Abundance
35.80	595.0
36.90	3062.0
37.90	2817.0
38.90	1243.0
39.80	207.0
43.80	834.0
44.80	587.0
45.80	107.0
46.90	984.0
47.70	540.0
48.90	2884.0
49.90	13281.0
50.90	4112.0
51.90	196.0
54.90	194.0
55.90	1055.0
56.90	1870.0
57.80	106.0
59.90	726.0
60.90	3099.0
61.90	3347.0
62.90	2783.0
63.80	389.0
64.90	696.0
66.70	202.0
67.90	7234.0
68.90	7271.0
69.90	540.0
71.80	394.0
72.90	3175.0
73.90	12001.0
74.90	36920.0
76.00	3071.0
76.90	636.0
77.70	270.0
78.80	1819.0
79.80	686.0
80.80	1889.0
81.80	498.0
85.90	127.0
86.90	3597.0
87.90	3811.0
90.70	276.0
91.90	2180.0
92.90	3323.0
93.90	8819.0
94.90	77464.0
95.90	5161.0
97.00	159.0
103.70	243.0
104.80	127.0
105.80	283.0
115.80	261.0
116.80	515.0
117.70	293.0
118.80	436.0
127.80	237.0
129.70	208.0
134.80	131.0
136.70	112.0
140.80	643.0
142.00	101.0
142.90	800.0
145.00	169.0
145.80	108.0
147.70	202.0
154.80	165.0
156.70	121.0
160.80	111.0
171.80	1008.0
172.90	635.0
173.80	64672.0
174.90	4463.0
175.80	62192.0
176.90	3939.0
177.80	135.0

Instrument :
MSVOA_V
ClientSampleId :
BFB74