

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073119\
 Data File : VV012080.D
 Acq On : 30 Jul 2019 20:10
 Operator : SY/MD
 Sample : VV0731WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK60

Quant Time: Jul 31 05:52:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR073119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 31 05:31:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	116483	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	112382	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	55161	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26257	4.80	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.00%
7) Chloroethane-d5	1.59	69	20896	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.20%
11) 1,1-Dichloroethene-d2	2.13	63	43054	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.97	46	76272	50.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.48%
24) Chloroform-d	4.40	84	70348	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
26) 1,2-Dichloroethane-d4	5.08	65	41619	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	5.10	84	142316	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	6.12	67	39522	5.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.40%
41) Toluene-d8	7.36	98	131797	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.60%
43) trans-1,3-Dichloropropene-	7.67	79	16583	4.77	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.40%
46) 2-Hexanone-d5	8.14	63	61016	51.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.80%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	27807	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	54804	5.40	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073119\
Data File : VV012080.D
Acq On : 30 Jul 2019 20:10
Operator : SY/MD
Sample : VV0731WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK60

Quant Time: Jul 31 05:52:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR073119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 31 05:31:06 2019
Response via : Initial Calibration

