

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV073120\
 Data File : VV017779.D
 Acq On : 31 Jul 2020 17:10
 Operator : SY/MD
 Sample : L3520-03
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 D0AB7

Quant Time: Aug 01 05:55:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 01 04:42:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	142664	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	133421	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	68228	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	28345	4.16	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.20%
7) Chloroethane-d5	1.58	69	26036	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.12	63	48047	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	3.93	46	65843	40.40	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.80%
24) Chloroform-d	4.37	84	68339	4.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.80%
26) 1,2-Dichloroethane-d4	5.06	65	42531	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.07	84	120783	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.20%
36) 1,2-Dichloropropane-d6	6.09	67	33781	4.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.60%
41) Toluene-d8	7.34	98	110877	3.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.60%
45) trans-1,3-Dichloropropene-	7.64	79	14980	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
48) 2-Hexanone-d5	8.11	63	41943	36.53	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.06%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	22783	4.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.20%
65) 1,2-Dichlorobenzene-d4	11.65	152	49503	4.80	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.00%


Target Compounds					Ovalue
13) Acetone	2.20	43	17311	13.569 ug/L	# 58
16) Methylene chloride	2.52	84	2383	0.279 ug/L	96
25) Chloroform	4.41	83	9054	0.458 ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV073120\
Data File : VV017779.D
Acq On : 31 Jul 2020 17:10
Operator : SY/MD
Sample : L3520-03
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
D0AB7

Quant Time: Aug 01 05:55:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR072320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 01 04:42:38 2020
Response via : Initial Calibration





#13

Acetone

Concen: 13.569 ug/L

RT: 2.20 min Scan# 346

Delta R.T. -0.03 min

Lab File: VV017779.D

Acq: 31 Jul 2020 17:10

Instrument :

MSVOA_V

ClientSampleId :


D0AB7

Tot Ion: 43 Resp: 17311

Ion Ratio Lower Upper




43 100

58 24.6 0.0 18.6#



Abundance Ion 43.05 (42.75 to 43.75): VV0

Ion 58.05 (57.75 to 58.75): VV0



#16

Methylene chloride

Concen: 0.279 ug/L

RT: 2.52 min Scan# 446

Delta R.T. 0.00 min

Lab File: VV017779.D

Acq: 31 Jul 2020 17:10


Tgt Ion: 84 Resp: 2383

Ion Ratio Lower Upper

84 100

86 59.2 46.5 86.3



49 111.4 79.2 147.2




Abundance Ion 84.05 (83.75 to 84.75): VV0

Ion 86.00 (85.70 to 86.70): VV0

Ion 49.10 (48.80 to 49.80): VV0





#25
 Chloroform
 Concen: 0.458 ug/L
 RT: 4.41 min Scan# 1031
 Delta R.T. 0.01 min
 Lab File: VV017779.D
 Acq: 31 Jul 2020 17:10

Instrument :
 MSVOA_V
 ClientSampleId :
 D0AB7

Tot Ion: 83 Resp: 9054
 Ion Ratio Lower Upper
 83 100
 85 62.5 47.3 87.9

