

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080218\
 Data File : VV006809.D
 Acq On : 02 Aug 2018 15:52
 Operator : SY/MD
 Sample : J4225-21
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWP9

Quant Time: Aug 03 02:37:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 03 01:56:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	273941	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	238853	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	99981	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	92749	4.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.80%
7) Chloroethane-d5	1.58	69	81769	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.13	63	127027	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	3.97	46	252794	54.94	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.88%
24) Chloroform-d	4.41	84	159139	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.09	65	85955	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.10	84	328355	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.12	67	109932	5.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.20%
41) Toluene-d8	7.37	98	274866	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
43) trans-1,3-Dichloropropene-	7.67	79	36616	4.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.20%
46) 2-Hexanone-d5	8.14	63	182638	54.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.96%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	83286	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	93390	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds					Ovalue
13) Acetone	2.20	43	10222	3.15 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080218\
Data File : VV006809.D
Acq On : 02 Aug 2018 15:52
Operator : SY/MD
Sample : J4225-21
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWP9

Quant Time: Aug 03 02:37:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 03 01:56:09 2018
Response via : Initial Calibration

