

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080218\
 Data File : VV006810.D
 Acq On : 02 Aug 2018 16:19
 Operator : SY/MD
 Sample : J4225-12DL 5X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWR0DL

Quant Time: Aug 03 10:59:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 03 01:56:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	240095	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	215606	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	87410	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87810	5.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.60%
7) Chloroethane-d5	1.58	69	77296	5.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.00%
11) 1,1-Dichloroethene-d2	2.13	63	119630	3.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.60%
20) 2-Butanone-d5	3.97	46	246384	61.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.18%
24) Chloroform-d	4.41	84	157865	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
26) 1,2-Dichloroethane-d4	5.09	65	83623	5.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.00%
32) Benzene-d6	5.10	84	316046	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
36) 1,2-Dichloropropane-d6	6.12	67	103625	5.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.80%
41) Toluene-d8	7.37	98	258558	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
43) trans-1,3-Dichloropropene-	7.67	79	34693	5.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.00%
46) 2-Hexanone-d5	8.14	63	173313	57.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.54%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	80502	5.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	89505	5.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.40%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	22427	1.04	ug/L 89
13) Acetone	2.20	43	5752	2.02	ug/L 96
18) trans-1,2-Dichloroethene	2.79	96	6199	0.37	ug/L 95
22) cis-1,2-Dichloroethene	3.96	96	254423	14.18	ug/L 97
34) Trichloroethene	5.96	95	8003	0.47	ug/L 93
47) Tetrachloroethene	8.03	164	3434	0.26	ug/L 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080218\
Data File : VV006810.D
Acq On : 02 Aug 2018 16:19
Operator : SY/MD
Sample : J4225-12DL 5X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWR0DL

Quant Time: Aug 03 10:59:15 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 03 01:56:09 2018
Response via : Initial Calibration

