

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080218\
 Data File : VV006841.D
 Acq On : 03 Aug 2018 07:06
 Operator : SY/MD
 Sample : J4227-09
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWR8

Quant Time: Aug 03 11:53:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 03 04:21:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	223822	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	202494	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	77820	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80514	5.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.80%
7) Chloroethane-d5	1.58	69	72939	5.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.20%
11) 1,1-Dichloroethene-d2	2.13	63	111942	3.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.00%
20) 2-Butanone-d5	3.95	46	226389	60.21	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.42%
24) Chloroform-d	4.41	84	148474	5.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.40%
26) 1,2-Dichloroethane-d4	5.09	65	78984	5.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.40%
32) Benzene-d6	5.10	84	283737	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
36) 1,2-Dichloropropane-d6	6.12	67	98090	5.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.80%
41) Toluene-d8	7.37	98	220111	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	7.67	79	27044	4.24	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.80%
46) 2-Hexanone-d5	8.14	63	141220	49.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.38%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	77933	5.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	116.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	76747	5.32	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080218\
Data File : VV006841.D
Acq On : 03 Aug 2018 07:06
Operator : SY/MD
Sample : J4227-09
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWR8

Quant Time: Aug 03 11:53:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 03 04:21:42 2018
Response via : Initial Calibration

