

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080219\
 Data File : VV012093.D
 Acq On : 02 Aug 2019 09:37
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00557

Quant Time: Aug 07 11:37:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR072919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 02 23:54:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	5.000	5.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	5.000	4.871	2.6	95	0.00
3 T	Chloromethane	5.000	4.462	10.8	90	0.00
4 S	Vinyl Chloride-d3	5.000	4.563	8.7	88	0.00
5 T	Vinyl chloride	5.000	4.237	15.3	85	0.00
6 T	Bromomethane	5.000	4.019	19.6	79	0.00
7 S	Chloroethane-d5	5.000	4.885	2.3	91	0.00
8 T	Chloroethane	5.000	4.457	10.9	91	0.00
9 T	Trichlorofluoromethane	5.000	5.066	-1.3	102	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	5.000	5.249	-5.0	109	0.00
11 S	1,1-Dichloroethene-d2	5.000	5.040	-0.8	99	0.00
12 T	1,1-Dichloroethene	5.000	5.209	-4.2	109	0.00
13 T	Acetone	50.000	46.648	6.7	98	0.00
14 T	Carbon disulfide	5.000	4.850	3.0	100	0.00
15 T	Methyl Acetate	5.000	4.550	9.0	100	0.00
16 T	Methylene chloride	5.000	5.034	-0.7	106	0.00
17 T	Methyl tert-butyl Ether	5.000	5.427	-8.5	110	0.00
18 T	trans-1,2-Dichloroethene	5.000	5.096	-1.9	105	0.00
19 T	1,1-Dichloroethane	5.000	4.879	2.4	98	0.00
20 S	2-Butanone-d5	50.000	47.973	4.1	88	0.00
21 T	2-Butanone	50.000	47.730	4.5	96	0.00
22 T	cis-1,2-Dichloroethene	5.000	5.157	-3.1	106	0.00
23 T	Bromochloromethane	5.000	5.593	-11.9	113	0.00
24 S	Chloroform-d	5.000	5.742	-14.8	110	0.00
25 T	Chloroform	5.000	4.617	7.7	93	0.00
26 S	1,2-Dichloroethane-d4	5.000	5.585	-11.7	103	0.00
27 T	1,2-Dichloroethane	5.000	5.451	-9.0	112	0.00
28 I	Chlorobenzene-d5	5.000	5.000	0.0	97	0.00
29 T	1,1,1-Trichloroethane	5.000	5.629	-12.6	110	0.00
30 T	Cyclohexane	5.000	4.739	5.2	92	0.00
31 T	Carbon tetrachloride	5.000	5.859	-17.2	113	0.00
32 S	Benzene-d6	5.000	5.192	-3.8	96	0.00
33 T	Benzene	5.000	5.070	-1.4	98	0.00
34 T	Trichloroethene	5.000	5.000	0.0	99	0.00
35 T	Methylcyclohexane	5.000	5.084	-1.7	100	0.00
36 S	1,2-Dichloropropane-d6	5.000	4.892	2.2	90	0.00
37 T	1,2-Dichloropropane	5.000	5.050	-1.0	97	0.00
38 T	Bromodichloromethane	5.000	5.509	-10.2	106	0.00
39 T	cis-1,3-Dichloropropene	5.000	5.216	-4.3	99	0.00
40 T	4-Methyl-2-pentanone	50.000	47.261	5.5	90	0.00
41 S	Toluene-d8	5.000	5.366	-7.3	96	0.00
42 T	Toluene	5.000	5.198	-4.0	98	0.00
43 S	trans-1,3-Dichloropropene-d	5.000	5.564	-11.3	100	0.00
44 T	trans-1,3-Dichloropropene	5.000	5.343	-6.9	102	0.00
45 T	1,1,2-Trichloroethane	5.000	4.931	1.4	96	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080219\
 Data File : VV012093.D
 Acq On : 02 Aug 2019 09:37
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00557

Quant Time: Aug 07 11:37:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR072919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 02 23:54:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	50.000	48.736	2.5	88	0.00
47 T	Tetrachloroethene	5.000	5.531	-10.6	103	0.00
48 T	2-Hexanone	50.000	48.807	2.4	92	0.00
49 T	Dibromochloromethane	5.000	5.910	-18.2	112	0.00
50 T	1,2-Dibromoethane	5.000	5.234	-4.7	101	0.00
51 T	Chlorobenzene	5.000	5.116	-2.3	100	0.00
52 T	Ethylbenzene	5.000	5.164	-3.3	99	0.00
53 T	m,p-xylene	5.000	5.302	-6.0	98	0.00
54 T	o-xylene	5.000	5.340	-6.8	102	0.00
55 T	Styrene	5.000	5.356	-7.1	100	0.00
56 S	1,1,2,2-Tetrachloroethane-d	5.000	5.043	-0.9	93	0.00
57 T	1,1,2,2-Tetrachloroethane	5.000	5.075	-1.5	98	0.00
58 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	96	0.00
59 T	Bromoform	5.000	5.984	-19.7	121	0.00
60 T	Isopropylbenzene	5.000	5.256	-5.1	103	0.00
61 T	1,2,3-Trichloropropane	5.000	4.779	4.4	95	0.00
62 T	1,3,5-Trimethylbenzene	5.000	5.495	-9.9	106	0.00
63 T	1,2,4-Trimethylbenzene	5.000	5.476	-9.5	103	0.00
64 T	1,3-Dichlorobenzene	5.000	5.302	-6.0	106	0.00
65 T	1,4-Dichlorobenzene	5.000	5.041	-0.8	100	0.00
66 S	1,2-Dichlorobenzene-d4	5.000	5.351	-7.0	98	0.00
67 T	1,2-Dichlorobenzene	5.000	5.179	-3.6	103	0.00
68 T	1,2-Dibromo-3-chloropropane	5.000	6.460	-29.2	125	0.00
69	1,3,5-Trichlorobenzene	5.000	5.532	-10.6	113	0.00
70 T	1,2,4-trichlorobenzene	5.000	5.727	-14.5	117	0.00
71	Naphthalene	5.000	5.685	-13.7	119	0.00
72 T	1,2,3-Trichlorobenzene	5.000	5.707	-14.1	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0