

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080318\
 Data File : VV006866.D
 Acq On : 04 Aug 2018 00:43
 Operator : SY/MD
 Sample : J4314-12
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KX0

Quant Time: Aug 04 05:51:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 04 04:38:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	198913	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	188522	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	71772	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	91326	6.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	132.60%#
7) Chloroethane-d5	1.58	69	81523	6.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	136.20%#
11) 1,1-Dichloroethene-d2	2.13	63	130419	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.20%
20) 2-Butanone-d5	3.97	46	234940	70.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	140.62%#
24) Chloroform-d	4.41	84	158353	6.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	131.40%#
26) 1,2-Dichloroethane-d4	5.09	65	83455	6.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	137.20%#
32) Benzene-d6	5.10	84	294950	5.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.80%
36) 1,2-Dichloropropane-d6	6.12	67	101232	6.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	120.60%
41) Toluene-d8	7.37	98	224600	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
43) trans-1,3-Dichloropropene-	7.67	79	31301	5.27	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.40%
46) 2-Hexanone-d5	8.15	63	135863	51.34	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.68%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	82562	6.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	132.00%#
64) 1,2-Dichlorobenzene-d4	11.68	152	77142	5.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.80%

Target Compounds					Ovalue
13) Acetone	2.20	43	7385	3.13 ug/L	94
17) Methyl tert-butyl Ether	2.82	73	17538	0.51 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080318\
Data File : VV006866.D
Acq On : 04 Aug 2018 00:43
Operator : SY/MD
Sample : J4314-12
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KX0

Quant Time: Aug 04 05:51:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 04 04:38:24 2018
Response via : Initial Calibration

