

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080318\
 Data File : VV006854.D
 Acq On : 03 Aug 2018 19:22
 Operator : SY/MD
 Sample : VV0803WBL01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK78

Quant Time: Aug 04 04:43:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 04 04:38:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	215998	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	204060	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	80332	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	86020	5.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.00%
7) Chloroethane-d5	1.58	69	76367	5.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.40%
11) 1,1-Dichloroethene-d2	2.13	63	121952	4.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.00%
20) 2-Butanone-d5	3.97	46	223150	61.50	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.00%
24) Chloroform-d	4.41	84	150739	5.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.20%
26) 1,2-Dichloroethane-d4	5.09	65	79358	6.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	120.20%
32) Benzene-d6	5.10	84	290673	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.40%
36) 1,2-Dichloropropane-d6	6.12	67	97854	5.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.60%
41) Toluene-d8	7.37	98	225830	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
43) trans-1,3-Dichloropropene-	7.67	79	31253	4.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.20%
46) 2-Hexanone-d5	8.15	63	140209	48.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.90%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	79653	5.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	78481	5.27	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080318\
Data File : VV006854.D
Acq On : 03 Aug 2018 19:22
Operator : SY/MD
Sample : VV0803WBL01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK78

Quant Time: Aug 04 04:43:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 04 04:38:24 2018
Response via : Initial Calibration

