

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080418\
 Data File : VV006882.D
 Acq On : 04 Aug 2018 15:12
 Operator : SY/MD
 Sample : J4314-13
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KX1

Quant Time: Aug 06 03:52:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Aug 06 03:23:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	252165	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	240489	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	91928	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100216	5.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.80%
7) Chloroethane-d5	1.58	69	92014	6.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.20%
11) 1,1-Dichloroethene-d2	2.13	63	144239	4.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.20%
20) 2-Butanone-d5	3.96	46	242850	57.33	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.66%
24) Chloroform-d	4.41	84	180767	5.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.40%
26) 1,2-Dichloroethane-d4	5.09	65	92343	5.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.80%
32) Benzene-d6	5.10	84	341871	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
36) 1,2-Dichloropropane-d6	6.13	67	118543	5.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.60%
41) Toluene-d8	7.37	98	258552	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
43) trans-1,3-Dichloropropene-	7.67	79	36403	4.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.00%
46) 2-Hexanone-d5	8.14	63	141060	41.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.58%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	89697	5.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	89506	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds					Ovalue
13) Acetone	2.20	43	8984	3.01 ug/L	89
34) Trichloroethene	5.96	95	15823	0.84 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080418\
Data File : VV006882.D
Acq On : 04 Aug 2018 15:12
Operator : SY/MD
Sample : J4314-13
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KX1

Quant Time: Aug 06 03:52:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080218WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Aug 06 03:23:10 2018
Response via : Initial Calibration

