

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080421\
 Data File : VV021733.D
 Acq On : 04 Aug 2021 11:24
 Operator : SY/MD
 Sample : VSTD00115
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001215

Quant Time: Aug 05 02:31:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM080421.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 05 02:29:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.639	114	14115	0.50 ug/L	0.00
5) Chlorobenzene-d5	8.863	117	12988	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6422	0.50 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.303	65	7639	0.90 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	7709	1.07 ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	10345	1.04 ug/L	0.00
8) Toluene-d8	7.317	98	29296	1.04 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	6214	1.13 ug/L	0.00
Target Compounds					
					Qvalue
3) Vinyl chloride	1.303	62	13586	1.13 ug/L	99
6) Trichloroethene	5.925	95	11059	1.08 ug/L	99
9) 1,2-Dibromoethane	8.368	107	7009	1.07 ug/L	99
12) 1,2,3-Trichloropropane	10.267	75	5819	1.16 ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.426	75	787	0.98 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080421\
Data File : VV021733.D
Acq On : 04 Aug 2021 11:24
Operator : SY/MD
Sample : VSTD00115
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001215

Quant Time: Aug 05 02:31:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM080421.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Aug 05 02:29:54 2021
Response via : Initial Calibration

