

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM080421.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Aug 06 01:25:29 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VV021732.D 0.05=VV021730.D 0.1 =VV021731.D 1.0 =VV021733.D 2 =VV021734.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.264	0.262	0.231	0.271	0.254	0.257	6.01
3) T	Vinyl chloride	0.475	0.461	0.459	0.481	0.458	0.467	2.25
4) S	1,2-Dichloroetha...	0.266	0.239	0.238	0.273	0.265	0.256	6.48
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.418	0.376	0.392	0.426	0.411	0.404	5.04
7) S	1,2-Dichloroprop...	0.399	0.331	0.330	0.398	0.378	0.367	9.41
8) S	Toluene-d8	1.110	1.053	0.942	1.128	1.062	1.059	6.87
9) T	1,2-Dibromoethane	0.263	0.231	0.233	0.270	0.258	0.251	7.16
10) S	1,1,2,2-Tetrachl...	0.226	0.196	0.192	0.239	0.229	0.216	9.73
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.430	0.387	0.407	0.453	0.440	0.423	6.23
13) T	1,2-Dibromo-3-ch...	0.056	0.065	0.071	0.061	0.060	0.063	8.74

(#) = Out of Range