

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080521\
 Data File : VV021740.D
 Acq On : 05 Aug 2021 12:04
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5284

Quant Time: Aug 06 01:11:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM080421.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 05 02:34:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	18765	0.50	ug/L	-0.02
5) Chlorobenzene-d5	8.864	117	16511	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	8267	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	4969	0.52	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	104.00%
4) 1,2-Dichloroethane-d4	5.042	65	4982	0.52	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
7) 1,2-Dichloropropane-d6	6.069	67	6560	0.54	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.00%
8) Toluene-d8	7.317	98	18668	0.53	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	3567	0.50	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	100.00%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.303	62	8475	0.48	ug/L	99
6) Trichloroethene	5.925	95	6925	0.52	ug/L	99
9) 1,2-Dibromoethane	8.353	107	4277	0.52	ug/L	95
12) 1,2,3-Trichloropropane	10.267	75	3492	0.50	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.427	75	453	0.44	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080521\
Data File : VV021740.D
Acq On : 05 Aug 2021 12:04
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5284

Quant Time: Aug 06 01:11:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM080421.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Aug 05 02:34:55 2021
Response via : Initial Calibration

