

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080522\
 Data File : VW027222.D
 Acq On : 05 Aug 2022 13:36
 Operator : SY/MD
 Sample : VIBLK271
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK271

Quant Time: Aug 06 05:17:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 06 03:57:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	356340	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	361419	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	141601	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	163110	67.139	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 134.280%		
7) Chloroethane-d5	1.561	69	114920	64.705	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 129.400%		
11) 1,1-Dichloroethene-d2	2.102	63	213998	48.156	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 96.320%		
21) 2-Butanone-d5	3.870	46	220545	101.986	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 101.990%		
24) Chloroform-d	4.336	84	285275	57.325	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 114.640%		
26) 1,2-Dichloroethane-d4	5.021	65	191745	62.058	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 124.120%		
32) Benzene-d6	5.040	84	571637	65.955	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 131.920%#		
36) 1,2-Dichloropropane-d6	6.063	67	187997	60.978	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 121.960%#		
41) Toluene-d8	7.310	98	465257	62.978	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 125.960%#		
43) trans-1,3-Dichloroprop...	7.616	79	76686	57.066	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 114.140%		
47) 2-Hexanone-d5	8.085	63	139470	93.925	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 93.930%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	245180	48.800	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 97.600%		
66) 1,2-Dichlorobenzene-d4	11.622	152	160546	65.654	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 131.300%#		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080522\
Data File : VV027222.D
Acq On : 05 Aug 2022 13:36
Operator : SY/MD
Sample : VIBLK271
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK271

Quant Time: Aug 06 05:17:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 06 03:57:43 2022
Response via : Initial Calibration

