

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080619\
 Data File : VV012113.D
 Acq On : 06 Aug 2019 14:44
 Operator : SY/MD
 Sample : VV0806MBL01
 Misc : 5.00µ/5.0mL/100µL/5mL/MSVOA_V/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK333

Quant Time: Aug 06 18:31:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM080619W.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 06 18:20:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	274890	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	249658	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.29	152	120970	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	70453	46.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.78%
7) Chloroethane-d5	1.58	69	51300	48.10	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.20%
11) 1,1-Dichloroethene-d2	2.13	63	103916	39.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.78%
21) 2-Butanone-d5	3.92	46	119005	108.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.70%
24) Chloroform-d	4.40	84	192117	50.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.92%
26) 1,2-Dichloroethane-d4	5.08	65	126861	52.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.76%
32) Benzene-d6	5.10	84	383021	53.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.62%
36) 1,2-Dichloropropane-d6	6.12	67	109789	52.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.50%
41) Toluene-d8	7.36	98	365436	52.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.50%
43) trans-1,3-Dichloropropene-	7.66	79	61894	52.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.62%
47) 2-Hexanone-d5	8.13	63	95000	109.91	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.91%
56) 1,1,2,2-Tetrachloroethane-	10.26	84	154025	52.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.02%
66) 1,2-Dichlorobenzene-d4	11.67	152	156061	54.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.70%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV080619\
Data File : VV012113.D
Acq On : 06 Aug 2019 14:44
Operator : SY/MD
Sample : VV0806MBL01
Misc : 5.00µ/5.0mL/100µL/5mL/MSVOA_V/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK333

Quant Time: Aug 06 18:31:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM080619W.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 06 18:20:54 2019
Response via : Initial Calibration

