

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080620\
 Data File : VV017813.D
 Acq On : 06 Aug 2020 14:50
 Operator : SY/MD
 Sample : L3590-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F2L52

Quant Time: Aug 07 06:23:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 07 05:27:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	218790	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	213488	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	117384	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	41514	38.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.82%
7) Chloroethane-d5	1.58	69	38817	42.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.22%
11) 1,1-Dichloroethene-d2	2.12	63	78766	30.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.44%
21) 2-Butanone-d5	3.91	46	82267	102.54	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.54%
24) Chloroform-d	4.38	84	127697	43.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.06	65	102339	48.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.10%
32) Benzene-d6	5.07	84	214399	42.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.90%
36) 1,2-Dichloropropane-d6	6.09	67	55921	40.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.18%
41) Toluene-d8	7.34	98	214791	43.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.74%
43) trans-1,3-Dichloropropene-	7.64	79	40312	43.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.46%
47) 2-Hexanone-d5	8.11	63	61161	102.01	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.01%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	89442	42.46	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.92%
64) 1,2-Dichlorobenzene-d4	11.65	152	114687	49.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.54%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	197912	127.512	ug/L 99
17) trans-1,2-Dichloroethene	2.78	96	11841	7.975	ug/L 89
20) cis-1,2-Dichloroethene	3.94	96	419405	264.450	ug/L 95
34) Trichloroethene	5.94	95	18880	11.472	ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080620\
Data File : VV017813.D
Acq On : 06 Aug 2020 14:50
Operator : SY/MD
Sample : L3590-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F2L52

Quant Time: Aug 07 06:23:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 07 05:27:52 2020
Response via : Initial Calibration

