

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080718\
 Data File : VV006914.D
 Acq On : 07 Aug 2018 23:03
 Operator : SY/MD
 Sample : J4357-03
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWS2

Quant Time: Aug 08 07:36:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 08 07:06:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	185533	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	177110	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	66190	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	66497	5.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	119.60%
7) Chloroethane-d5	1.58	69	58318	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.20%
11) 1,1-Dichloroethene-d2	2.13	63	94021	4.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.20%
20) 2-Butanone-d5	3.97	46	199093	56.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.18%
24) Chloroform-d	4.41	84	127493	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.20%
26) 1,2-Dichloroethane-d4	5.09	65	69452	5.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.60%
32) Benzene-d6	5.10	84	236912	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.20%
36) 1,2-Dichloropropane-d6	6.12	67	82910	5.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.20%
41) Toluene-d8	7.37	98	178737	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
43) trans-1,3-Dichloropropene-	7.67	79	22333	4.39	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.80%
46) 2-Hexanone-d5	8.15	63	98797	46.71	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.42%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	62849	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	61326	5.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.60%

Target Compounds					Ovalue
13) Acetone	2.21	43	15449	6.67 ug/L	100
22) cis-1,2-Dichloroethene	3.97	96	14315	1.07 ug/L	96
27) 1,2-Dichloroethane	5.19	62	10732	0.67 ug/L	97
34) Trichloroethene	5.96	95	3444	0.27 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080718\
Data File : VV006914.D
Acq On : 07 Aug 2018 23:03
Operator : SY/MD
Sample : J4357-03
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWS2

Quant Time: Aug 08 07:36:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 08 07:06:57 2018
Response via : Initial Calibration

