

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080718\
 Data File : VV006916.D
 Acq On : 07 Aug 2018 23:57
 Operator : SY/MD
 Sample : J4357-05
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWS4

Quant Time: Aug 08 07:40:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 08 07:06:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	183997	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	172164	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	64525	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	66925	6.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	121.40%
7) Chloroethane-d5	1.58	69	60415	5.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.00%
11) 1,1-Dichloroethene-d2	2.13	63	93731	4.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.60%
20) 2-Butanone-d5	3.97	46	207861	59.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.10%
24) Chloroform-d	4.40	84	131091	5.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.20%
26) 1,2-Dichloroethane-d4	5.09	65	73587	5.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.20%
32) Benzene-d6	5.10	84	248443	5.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.80%
36) 1,2-Dichloropropane-d6	6.12	67	88577	5.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.60%
41) Toluene-d8	7.37	98	182037	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
43) trans-1,3-Dichloropropene-	7.67	79	23444	4.74	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.80%
46) 2-Hexanone-d5	8.15	63	104425	50.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.58%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	66542	5.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	62465	5.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.60%

Target Compounds						Ovalue
5) Vinyl chloride	1.32	62	27309	1.58	ug/L	87
13) Acetone	2.21	43	10954	4.77	ug/L	99
53) m,p-xylene	9.19	106	4193	0.24	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080718\
Data File : VV006916.D
Acq On : 07 Aug 2018 23:57
Operator : SY/MD
Sample : J4357-05
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWS4

Quant Time: Aug 08 07:40:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 08 07:06:57 2018
Response via : Initial Calibration

