

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080718\
 Data File : VV006926.D
 Acq On : 08 Aug 2018 04:24
 Operator : SY/MD
 Sample : J4357-12MS
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWT0MS

Quant Time: Aug 08 07:58:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 08 07:06:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	208945	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	196537	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	76396	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	81256	6.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	129.80%
7) Chloroethane-d5	1.58	69	63714	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.00%
11) 1,1-Dichloroethene-d2	2.13	63	132463	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.80%
20) 2-Butanone-d5	3.97	46	236713	59.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.44%
24) Chloroform-d	4.41	84	141058	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.20%
26) 1,2-Dichloroethane-d4	5.09	65	78535	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.00%
32) Benzene-d6	5.10	84	273824	5.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.80%
36) 1,2-Dichloropropane-d6	6.12	67	96708	5.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.60%
41) Toluene-d8	7.37	98	207429	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
43) trans-1,3-Dichloropropene-	7.67	79	26146	4.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.60%
46) 2-Hexanone-d5	8.15	63	122410	52.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.32%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	74000	5.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	75912	6.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	122.80%#

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.32	62	1215922	61.91	ug/L	98
12) 1,1-Dichloroethene	2.14	96	61588	4.46	ug/L	91
13) Acetone	2.21	43	7211	2.76	ug/L	100
22) cis-1,2-Dichloroethene	3.97	96	351729	23.42	ug/L	98
33) Benzene	5.15	78	290639	4.79	ug/L	100
34) Trichloroethene	5.96	95	75768	5.37	ug/L	97
42) Toluene	7.44	91	260815	4.59	ug/L	99
47) Tetrachloroethene	8.03	164	6835	0.64	ug/L	95
51) Chlorobenzene	8.93	112	177773	5.14	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080718\
Data File : VV006926.D
Acq On : 08 Aug 2018 04:24
Operator : SY/MD
Sample : J4357-12MS
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWT0MS

Quant Time: Aug 08 07:58:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 08 07:06:57 2018
Response via : Initial Calibration

