

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080718\
 Data File : VV006911.D
 Acq On : 07 Aug 2018 21:43
 Operator : SY/MD
 Sample : J4336-02
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Aug 08 07:29:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 08 07:06:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	185379	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	175244	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	68356	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54607	4.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.20%
7) Chloroethane-d5	1.58	69	50313	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	2.13	63	80790	3.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.80%
20) 2-Butanone-d5	3.98	46	172805	48.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.44%
24) Chloroform-d	4.41	84	114675	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.09	65	65179	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.80%
32) Benzene-d6	5.10	84	218104	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
36) 1,2-Dichloropropane-d6	6.12	67	78498	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.60%
41) Toluene-d8	7.37	98	163139	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.67	79	23939	4.76	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.20%
46) 2-Hexanone-d5	8.15	63	101926	48.71	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.42%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	62468	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	55877	5.05	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080718\
Data File : VV006911.D
Acq On : 07 Aug 2018 21:43
Operator : SY/MD
Sample : J4336-02
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Aug 08 07:29:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 08 07:06:57 2018
Response via : Initial Calibration

