

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080719\
 Data File : VV012121.D
 Acq On : 07 Aug 2019 12:31
 Operator : SY/MD
 Sample : VV0807WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK61

Quant Time: Aug 08 01:22:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM080619W.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 08 01:11:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	243037	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	239219	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.29	152	115970	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	63634	47.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.80%
7) Chloroethane-d5	1.58	69	42721	45.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.60%
11) 1,1-Dichloroethene-d2	2.13	63	82539	35.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.68%
21) 2-Butanone-d5	3.92	46	108336	111.92	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.92%
24) Chloroform-d	4.40	84	166194	49.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.72%
26) 1,2-Dichloroethane-d4	5.08	65	113588	53.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.10%
32) Benzene-d6	5.10	84	323227	46.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.90%
36) 1,2-Dichloropropane-d6	6.12	67	95917	48.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.20%
41) Toluene-d8	7.36	98	338854	51.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.10%
43) trans-1,3-Dichloropropene-	7.66	79	57684	50.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.76%
47) 2-Hexanone-d5	8.13	63	93511	112.91	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.91%
56) 1,1,2,2-Tetrachloroethane-	10.26	84	147580	52.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.02%
66) 1,2-Dichlorobenzene-d4	11.67	152	146129	53.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.16%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080719\
Data File : VV012121.D
Acq On : 07 Aug 2019 12:31
Operator : SY/MD
Sample : VV0807WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK61

Quant Time: Aug 08 01:22:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM080619W.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 08 01:11:59 2019
Response via : Initial Calibration

