

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080719\  
 Data File : VV012121.D  
 Acq On : 07 Aug 2019 12:31  
 Operator : SY/MD  
 Sample : VV0807WBL01  
 Misc : 5.0mL/MSVOA V/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK61

Quant Time: Aug 08 01:22:12 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SFAMVLM080619W.M  
 Quant Title : VOC Analysis  
 QLast Update : Thu Aug 08 01:11:59 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.66  | 114  | 243037   | 50.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.89  | 117  | 239219   | 50.00 | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4 | 11.29 | 152  | 115970   | 50.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |         |       |          |          |      |         |
|--------------------------------|---------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.32    | 65    | 63634    | 47.90    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 135 | Recovery | =    | 95.80%  |
| 7) Chloroethane-d5             | 1.58    | 69    | 42721    | 45.30    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 130 | Recovery | =    | 90.60%  |
| 11) 1,1-Dichloroethene-d2      | 2.13    | 63    | 82539    | 35.84    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 71.68%  |
| 21) 2-Butanone-d5              | 3.92    | 46    | 108336   | 111.92   | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 40 - 130 | Recovery | =    | 111.92% |
| 24) Chloroform-d               | 4.40    | 84    | 166194   | 49.86    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 99.72%  |
| 26) 1,2-Dichloroethane-d4      | 5.08    | 65    | 113588   | 53.05    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 106.10% |
| 32) Benzene-d6                 | 5.10    | 84    | 323227   | 46.95    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 93.90%  |
| 36) 1,2-Dichloropropane-d6     | 6.12    | 67    | 95917    | 48.10    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 120 | Recovery | =    | 96.20%  |
| 41) Toluene-d8                 | 7.36    | 98    | 338854   | 51.05    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 102.10% |
| 43) trans-1,3-Dichloropropene- | 7.66    | 79    | 57684    | 50.88    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 101.76% |
| 47) 2-Hexanone-d5              | 8.13    | 63    | 93511    | 112.91   | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 45 - 130 | Recovery | =    | 112.91% |
| 56) 1,1,2,2-Tetrachloroethane- | 10.26   | 84    | 147580   | 52.51    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 65 - 120 | Recovery | =    | 105.02% |
| 66) 1,2-Dichlorobenzene-d4     | 11.67   | 152   | 146129   | 53.58    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 107.16% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080719\  
Data File : VV012121.D  
Acq On : 07 Aug 2019 12:31  
Operator : SY/MD  
Sample : VV0807WBL01  
Misc : 5.0mL/MSVOA V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK61

Quant Time: Aug 08 01:22:12 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SFAMVLM080619W.M  
Quant Title : VOC Analysis  
QLast Update : Thu Aug 08 01:11:59 2019  
Response via : Initial Calibration

