

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080720\
 Data File : VV017839.D
 Acq On : 07 Aug 2020 14:53
 Operator : SY/MD
 Sample : L3590-11
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Aug 08 01:27:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 08 00:22:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	232951	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	228006	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	125411	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	39732	34.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.96%
7) Chloroethane-d5	1.58	69	40716	41.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.96%
11) 1,1-Dichloroethene-d2	2.12	63	84144	30.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.64%
21) 2-Butanone-d5	3.91	46	77415	90.63	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.63%
24) Chloroform-d	4.38	84	143431	46.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.06	65	112339	49.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.08%
32) Benzene-d6	5.07	84	235639	43.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.36%
36) 1,2-Dichloropropane-d6	6.09	67	64308	43.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.34%
41) Toluene-d8	7.34	98	239548	45.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.58%
43) trans-1,3-Dichloropropene-	7.65	79	20267	20.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	41.16%#
47) 2-Hexanone-d5	8.11	63	55756	87.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.07%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	100109	44.50	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.00%
64) 1,2-Dichlorobenzene-d4	11.65	152	128660	52.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.52%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	115121	69.662	ug/L 98
17) trans-1,2-Dichloroethene	2.78	96	5244	3.317	ug/L 89
20) cis-1,2-Dichloroethene	3.94	96	30052	17.797	ug/L 92
34) Trichloroethene	5.95	95	2891m	1.645	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080720\
Data File : VV017839.D
Acq On : 07 Aug 2020 14:53
Operator : SY/MD
Sample : L3590-11
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Aug 08 01:27:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 08 00:22:47 2020
Response via : Initial Calibration

