

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080720\
 Data File : VV017828.D
 Acq On : 07 Aug 2020 10:36
 Operator : SY/MD
 Sample : VV0807WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK70

Quant Time: Aug 08 00:37:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 08 00:22:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	235780	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	231714	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	132856	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	44904	39.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.12%
7) Chloroethane-d5	1.58	69	44241	45.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.14%
11) 1,1-Dichloroethene-d2	2.12	63	92808	33.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.18%
21) 2-Butanone-d5	3.91	46	83723	96.84	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.84%
24) Chloroform-d	4.37	84	150899	48.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.04%
26) 1,2-Dichloroethane-d4	5.06	65	118571	51.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.32%
32) Benzene-d6	5.08	84	247800	45.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.40%
36) 1,2-Dichloropropane-d6	6.09	67	66465	43.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.80%
41) Toluene-d8	7.34	98	254149	47.28	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.56%
43) trans-1,3-Dichloropropene-	7.64	79	43507	43.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.96%
47) 2-Hexanone-d5	8.11	63	55203	84.83	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.83%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	101971	44.61	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	130567	50.07	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080720\
Data File : VV017828.D
Acq On : 07 Aug 2020 10:36
Operator : SY/MD
Sample : VV0807WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK70

Quant Time: Aug 08 00:37:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 08 00:22:47 2020
Response via : Initial Calibration

