

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080720\
 Data File : VV017832.D
 Acq On : 07 Aug 2020 12:10
 Operator : SY/MD
 Sample : L3572-17
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Aug 08 01:02:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 08 00:22:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	225813	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	225194	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	127064	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	44802	40.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.38%
7) Chloroethane-d5	1.58	69	44116	46.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.84%
11) 1,1-Dichloroethene-d2	2.12	63	96207	36.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.72%
21) 2-Butanone-d5	3.91	46	77952	94.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.14%
24) Chloroform-d	4.38	84	147433	48.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.98%
26) 1,2-Dichloroethane-d4	5.06	65	116284	52.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.80%
32) Benzene-d6	5.08	84	252206	47.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.68%
36) 1,2-Dichloropropane-d6	6.09	67	67056	45.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.14%
41) Toluene-d8	7.34	98	252422	48.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.64%
43) trans-1,3-Dichloropropene-	7.64	79	43394	44.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.24%
47) 2-Hexanone-d5	8.11	63	55531	87.80	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.80%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	101624	45.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.48%
64) 1,2-Dichlorobenzene-d4	11.65	152	131833	52.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.72%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080720\
Data File : VV017832.D
Acq On : 07 Aug 2020 12:10
Operator : SY/MD
Sample : L3572-17
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Aug 08 01:02:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 08 00:22:47 2020
Response via : Initial Calibration

