

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080720\
 Data File : VV017834.D
 Acq On : 07 Aug 2020 12:56
 Operator : SY/MD
 Sample : L3590-16DL 5X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4L21DL

Quant Time: Aug 08 01:09:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 08 00:22:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	230066	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	222372	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	127055	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	44640	39.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.58%
7) Chloroethane-d5	1.58	69	43628	45.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.08%
11) 1,1-Dichloroethene-d2	2.12	63	92733	34.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.80%
21) 2-Butanone-d5	3.91	46	80595	95.54	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.54%
24) Chloroform-d	4.38	84	148035	48.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.56%
26) 1,2-Dichloroethane-d4	5.06	65	116229	51.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.80%
32) Benzene-d6	5.08	84	250139	47.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.08%
36) 1,2-Dichloropropane-d6	6.09	67	65962	45.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.80%
41) Toluene-d8	7.34	98	251841	48.82	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.64%
43) trans-1,3-Dichloropropene-	7.64	79	43361	45.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.32%
47) 2-Hexanone-d5	8.11	63	55842	89.42	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.42%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	100689	45.89	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.78%
64) 1,2-Dichlorobenzene-d4	11.65	152	128060	51.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.70%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	57874	35.460 ug/L	95
17) trans-1,2-Dichloroethene	2.78	96	6528	4.181 ug/L	93
20) cis-1,2-Dichloroethene	3.94	96	77548	46.500 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080720\
Data File : VV017834.D
Acq On : 07 Aug 2020 12:56
Operator : SY/MD
Sample : L3590-16DL 5X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4L21DL

Quant Time: Aug 08 01:09:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 08 00:22:47 2020
Response via : Initial Calibration

