

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080720\
 Data File : VV017842.D
 Acq On : 07 Aug 2020 16:03
 Operator : SY/MD
 Sample : VIBLK50
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK50

Quant Time: Aug 08 01:35:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 08 00:22:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	241259	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	238360	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	131973	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	65277	55.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	110.98%
7) Chloroethane-d5	1.58	69	55820	55.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.14%
11) 1,1-Dichloroethene-d2	2.12	63	121975	43.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.28%
21) 2-Butanone-d5	3.91	46	78346	88.56	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.56%
24) Chloroform-d	4.38	84	161253	50.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.30%
26) 1,2-Dichloroethane-d4	5.06	65	124494	53.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.02%
32) Benzene-d6	5.08	84	277373	49.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.38%
36) 1,2-Dichloropropane-d6	6.09	67	71755	46.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.14%
41) Toluene-d8	7.34	98	278790	50.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.84%
43) trans-1,3-Dichloropropene-	7.64	79	26566	25.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	51.62%#
47) 2-Hexanone-d5	8.11	63	47474	70.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	70.92%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	105150	44.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.42%
64) 1,2-Dichlorobenzene-d4	11.65	152	139032	53.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.34%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080720\
Data File : VV017842.D
Acq On : 07 Aug 2020 16:03
Operator : SY/MD
Sample : VIBLK50
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK50

Quant Time: Aug 08 01:35:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM073020WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 08 00:22:47 2020
Response via : Initial Calibration

