

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080818\
 Data File : VV006944.D
 Acq On : 08 Aug 2018 19:52
 Operator : SY/MD
 Sample : J4357-18DL 100X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWT6DL

Quant Time: Aug 09 08:12:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Aug 09 07:05:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	190481	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	184299	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	72547	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	51334	4.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.80%
7) Chloroethane-d5	1.58	69	53133	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.00%
11) 1,1-Dichloroethene-d2	2.13	63	78348	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	3.98	46	198777	54.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.10%
24) Chloroform-d	4.41	84	133484	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.20%
26) 1,2-Dichloroethane-d4	5.09	65	72038	5.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.80%
32) Benzene-d6	5.10	84	244084	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
36) 1,2-Dichloropropane-d6	6.12	67	88694	5.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.00%
41) Toluene-d8	7.37	98	176754	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	7.67	79	25956	4.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.00%
46) 2-Hexanone-d5	8.15	63	112618	51.17	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.34%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	69743	5.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	65537	5.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.60%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	8065	0.45	ug/L # 77
22) cis-1,2-Dichloroethene	3.97	96	65744	4.80	ug/L 96
34) Trichloroethene	5.96	95	27007	2.04	ug/L 97
47) Tetrachloroethene	8.02	164	71968	7.15	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080818\
Data File : VV006944.D
Acq On : 08 Aug 2018 19:52
Operator : SY/MD
Sample : J4357-18DL 100X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWT6DL

Quant Time: Aug 09 08:12:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Aug 09 07:05:51 2018
Response via : Initial Calibration

