

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080819\
 Data File : VV012142.D
 Acq On : 08 Aug 2019 23:57
 Operator : SY/MD
 Sample : K4184-06
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA3

Quant Time: Aug 09 07:12:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 09 06:44:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	100277	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	95068	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	51045	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	23071	5.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.20%
7) Chloroethane-d5	1.58	69	16563	5.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.80%
11) 1,1-Dichloroethene-d2	2.13	63	35907	4.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.20%
20) 2-Butanone-d5	3.95	46	12994	11.80	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	23.60%#
24) Chloroform-d	4.40	84	58432	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.08	65	36098	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.00%
32) Benzene-d6	5.10	84	121405	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
36) 1,2-Dichloropropane-d6	6.12	67	31525	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.20%
41) Toluene-d8	7.36	98	122016	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.00%
43) trans-1,3-Dichloropropene-	7.66	79	15861	4.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.40%
46) 2-Hexanone-d5	8.13	63	43102	42.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.08%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	25930	5.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	51731	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds					Ovalue
13) Acetone	2.19	43	2163	3.338 ug/L	100
25) Chloroform	4.43	83	10450	0.843 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080819\
Data File : VV012142.D
Acq On : 08 Aug 2019 23:57
Operator : SY/MD
Sample : K4184-06
Misc : 25mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA3

Quant Time: Aug 09 07:12:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 09 06:44:52 2019
Response via : Initial Calibration

