

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080820\
 Data File : VV017862.D
 Acq On : 08 Aug 2020 18:36
 Operator : SY/MD
 Sample : VV0808WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK72

Quant Time: Aug 10 05:17:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Aug 10 05:09:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	250679	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	235040	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	122344	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	65511	50.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.20%
7) Chloroethane-d5	1.58	69	55410	50.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.76%
11) 1,1-Dichloroethene-d2	2.12	63	118229	38.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.02%
21) 2-Butanone-d5	3.91	46	85984	106.07	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.07%
24) Chloroform-d	4.37	84	167697	49.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.05	65	127150	51.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.12%
32) Benzene-d6	5.07	84	288332	53.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.24%
36) 1,2-Dichloropropane-d6	6.09	67	74979	54.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.96%
41) Toluene-d8	7.33	98	283398	51.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.90%
43) trans-1,3-Dichloropropene-	7.64	79	42327	47.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.52%
47) 2-Hexanone-d5	8.11	63	51189	102.70	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.70%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	109089	50.85	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.70%
64) 1,2-Dichlorobenzene-d4	11.65	152	133653	53.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080820\
Data File : VV017862.D
Acq On : 08 Aug 2020 18:36
Operator : SY/MD
Sample : VV0808WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK72

Quant Time: Aug 10 05:17:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 10 05:09:16 2020
Response via : Initial Calibration

