

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080820\
 Data File : VV017870.D
 Acq On : 08 Aug 2020 21:42
 Operator : SY/MD
 Sample : L3608-04
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F2L57

Quant Time: Aug 10 06:31:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Aug 10 05:09:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	216907	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	208682	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	106382	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	59599	53.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	106.40%
7) Chloroethane-d5	1.57	69	51380	53.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.98%
11) 1,1-Dichloroethene-d2	2.12	63	105022	39.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.04%
21) 2-Butanone-d5	3.91	46	69877	99.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.62%
24) Chloroform-d	4.37	84	145687	49.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.05	65	114287	53.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.10%
32) Benzene-d6	5.07	84	251933	52.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.56%
36) 1,2-Dichloropropane-d6	6.09	67	65281	53.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.86%
41) Toluene-d8	7.34	98	243972	50.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.74%
43) trans-1,3-Dichloropropene-	7.64	79	35002	44.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.98%
47) 2-Hexanone-d5	8.11	63	42111	95.16	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.16%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	93034	48.85	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.70%
64) 1,2-Dichlorobenzene-d4	11.65	152	116914	53.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.82%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	30403	23.989 ug/L	93
17) trans-1,2-Dichloroethene	2.78	96	942	0.722 ug/L #	85
20) cis-1,2-Dichloroethene	3.93	96	85308	60.560 ug/L	100
34) Trichloroethene	5.94	95	37976	25.471 ug/L	97
46) Tetrachloroethene	8.00	164	14292	10.170 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080820\
Data File : VV017870.D
Acq On : 08 Aug 2020 21:42
Operator : SY/MD
Sample : L3608-04
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F2L57

Quant Time: Aug 10 06:31:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 10 05:09:16 2020
Response via : Initial Calibration

