

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080820\
 Data File : VV017881.D
 Acq On : 09 Aug 2020 01:59
 Operator : SY/MD
 Sample : L3608-17
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4L19

Quant Time: Aug 10 07:36:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Aug 10 05:09:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	220615	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	211420	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	108902	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	60764	53.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	106.66%
7) Chloroethane-d5	1.58	69	52857	54.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.22%
11) 1,1-Dichloroethene-d2	2.12	63	107894	39.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.82%
21) 2-Butanone-d5	3.91	46	71773	100.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.60%
24) Chloroform-d	4.38	84	150788	50.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.94%
26) 1,2-Dichloroethane-d4	5.05	65	119042	54.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.70%
32) Benzene-d6	5.07	84	261577	53.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.16%
36) 1,2-Dichloropropane-d6	6.09	67	69763	56.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.72%
41) Toluene-d8	7.34	98	260141	53.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.02%
43) trans-1,3-Dichloropropene-	7.64	79	37467	47.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.00%
47) 2-Hexanone-d5	8.11	63	42886	95.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.66%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	91318	47.32	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.64%
64) 1,2-Dichlorobenzene-d4	11.65	152	118707	53.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.94%


Target Compounds					Ovalue
13) Acetone	2.19	43	6058	6.525 ug/L	80




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080820\
Data File : VV017881.D
Acq On : 09 Aug 2020 01:59
Operator : SY/MD
Sample : L3608-17
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4L19

Quant Time: Aug 10 07:36:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM080820WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 10 05:09:16 2020
Response via : Initial Calibration





#13

Acetone

Concen: 6.525 ug/L

RT: 2.19 min Scan# 343

Delta R.T. -0.00 min

Lab File: VV017881.D

Acq: 09 Aug 2020 01:59

Instrument :

MSVOA_V

ClientSampleId :

F4L19

Tot Ion: 43 Resp: 6058

Ion Ratio Lower Upper

43 100

58 17.2 0.0 54.8

