

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV080918\
 Data File : VV006970.D
 Acq On : 10 Aug 2018 00:27
 Operator : SY/MD
 Sample : J4357-20
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWT9

Quant Time: Aug 10 04:19:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 10 02:34:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	251362	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	236628	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	95709	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82158	5.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.20%
7) Chloroethane-d5	1.58	69	60810	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.40%
11) 1,1-Dichloroethene-d2	2.13	63	101897	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	3.97	46	291299	62.43	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.86%
24) Chloroform-d	4.41	84	148444	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.20%
26) 1,2-Dichloroethane-d4	5.09	65	81304	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
32) Benzene-d6	5.10	84	286712	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	6.12	67	100537	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.36	98	212803	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.20%
43) trans-1,3-Dichloropropene-	7.67	79	30250	4.18	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.60%
46) 2-Hexanone-d5	8.15	63	134170	45.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.76%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	80092	4.94	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	82133	4.93	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.60%

Target Compounds

						Ovalue
5) Vinyl chloride	1.32	62	843920	34.66	ug/L	88
18) trans-1,2-Dichloroethene	2.79	96	106984	5.54	ug/L	99
22) cis-1,2-Dichloroethene	3.97	96	10218293	465.90	ug/L	100
30) Cyclohexane	4.72	56	7305	0.22	ug/L	98
34) Trichloroethene	5.96	95	5516	0.28	ug/L	94
47) Tetrachloroethene	8.02	164	6159	0.39	ug/L	93
53) m,p-xylene	9.19	106	7284	0.25	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV080918\
Data File : VV006970.D
Acq On : 10 Aug 2018 00:27
Operator : SY/MD
Sample : J4357-20
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWT9

Quant Time: Aug 10 04:19:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 10 02:34:57 2018
Response via : Initial Calibration

