

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_V\DATA\VV080918\
 Data File : VV006958.D
 Acq On : 09 Aug 2018 18:40
 Operator : SY/MD
 Sample : J4357-01DLRE 4X
 Misc : 25 mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESWS0DLRE

Quant Time: Aug 10 02:51:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 10 02:34:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	228396	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	215646	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	84132	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	71116	4.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.40%
7) Chloroethane-d5	1.58	69	63549	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.60%
11) 1,1-Dichloroethene-d2	2.13	63	96683	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	3.97	46	209183	49.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.68%
24) Chloroform-d	4.41	84	145824	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.09	65	78246	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.60%
32) Benzene-d6	5.10	84	277999	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
36) 1,2-Dichloropropane-d6	6.12	67	101198	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.36	98	207900	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
43) trans-1,3-Dichloropropene-	7.67	79	29961	4.55	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.00%
46) 2-Hexanone-d5	8.15	63	118045	43.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.62%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	72232	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	75240	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%


Target Compounds					Qvalue
5) Vinyl chloride	1.32	62	25540	1.155 ug/L	94
22) cis-1,2-Dichloroethene	3.97	96	111493	5.595 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_V\DATA\VV080918\
Data File : VV006958.D
Acq On : 09 Aug 2018 18:40
Operator : SY/MD
Sample : J4357-01DLRE 4X
Misc : 25 mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESWS0DLRE

Quant Time: Aug 10 02:51:21 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR080918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 10 02:34:57 2018
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 1.155 ug/L

RT: 1.32 min Scan# 134

Delta R.T. -0.00 min

Lab File: VV006958.D

Acq: 09 Aug 2018 18:40

Instrument :

MSVOA_V

ClientSampleId :





ESWS0DLRE

Tgt Ion: 62 Resp: 25540

Ion Ratio Lower Upper

62 100

64 37.2 23.5 43.7



#22

cis-1,2-Dichloroethene

Concen: 5.595 ug/L

RT: 3.97 min Scan# 957

Delta R.T. 0.00 min

Lab File: VV006958.D

Acq: 09 Aug 2018 18:40

Tgt Ion: 96 Resp: 111493

Ion Ratio Lower Upper

96 100

61 127.6 91.1 169.1

68 0.0 0.0 0.0

